



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis reagent for post-column analysis of Carbamate pesticides

Part Number: CB130

Lot Number: 109040

Production Date: 30 Sep 2011

Expiration Date: 29 Sep 2013

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) y
2. Baseline noise (0 ± 25 mvolts) (Y/N) y
3. Twelve appropriate peaks in chromatogram (Y/N) y

Sample: Carbamate Test Mixture, PN: 1700-0063.

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by:  **Date:** 05 Oct 2011

VP Oper
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. A, 04/08/2011

P:\Documentation\Chemical Documentation\CB130