



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis reagent for post-column analysis of Carbamate pesticides

Part Number: CB130

Lot Number: 110 004

Production Date: 06 Oct 2011

Expiration Date: 05 Oct 2013

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Baseline noise (0 ± 25 mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

Y
Y
Y

Sample: Carbamate Test Mixture, PN: 1700-0063.

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: _____

Jim P...
VP OPER
Title

Date: 07 Oct 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. A, 04/08/2011

P:\Documentation\Chemical Documentation\CB130