



## CERTIFICATE OF ANALYSIS

### Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis reagent for post-column analysis of Carbamate pesticides

Part Number: CB130

Lot Number: 111043

Production Date: 30 Nov 2011

Expiration Date: 30 Nov 2013

Primary Application: Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

1. pH =  $12.5 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

Y  
Y  
Y

Sample: Carbamate Test Mixture, PN: 1700-0063.

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: [Signature]

VP Dir  
Title

Date: 05 Dec 2011

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE