



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis reagent for post-column analysis of Carbamate pesticides

Part Number: CB130

Lot Number: 201022

Production Date: 18 JAN 2012

Expiration Date: 18 JAN 2014

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria


1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Baseline noise (0 ± 25 mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

Y
Y
Y

Sample: Carbamate Test Mixture, PN: 1700-0063.

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by:


Title

Date:

09 FEB 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE