



CERTIFICATE OF ANALYSIS

Hydrolysis Reagent for Post-column Carbamate Pesticides Analysis

CB130

Product: Hydrolysis reagent for post-column analysis of Carbamate pesticides

Part Number: CB130

Lot Number: 2509066

Production Date: 23 Sep 2025

Expiration Date: 23 Sep 2027

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

1. pH = 12.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)	<u>Y</u>
2. Baseline noise (0 ± 25 mvolts) (Y/N)	<u>Y</u>
3. Twelve appropriate peaks in chromatogram (Y/N)	<u>Y</u>

Sample: Carbamate Test Mixture, **PN:** 1700-0063.

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: David Mazawa **Date:** 1 Oct 2025
Title: QC Manager