



## CERTIFICATE OF ANALYSIS

### OPA Diluent for Post-Column Carbamate Analysis

**CB910**

**Product:** OPA Diluent for Post Column Carbamate Analysis

**Part Number:** CB910

**Lot Number:** 006004

**Test Start Date:** 03 Jun 2010

**Expiration Date:** 02 Jun 2012

**Primary Application:** For preparation of post-column fluorescence reagent for carbamate pesticide analysis.

#### Acceptance Criteria:

1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C 9.040
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y
3. Twelve appropriate peaks in chromatogram (Y/N) Y

**Calibration Standard:** Carbamate Test Standard, PN 1700-0063, LN 908045

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** [Signature] **Date:** 04 Jun 2010

Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 8, 9/19/08