



# CERTIFICATE OF ANALYSIS

## OPA Diluent for Post-Column Carbamate Analysis

CB910

Product: OPA Diluent for Post Column Carbamate Analysis

Part Number: CB910

Lot Number: 007036

Test Start Date: 28 Jul 2010

Expiration Date: 27 Jul 2012

**Primary Application:** For preparation of post-column fluorescence reagent for carbamate pesticide analysis.

### Acceptance Criteria:

- |   |              |
|---|--------------|
| 1. pH = $9.10 \pm 0.1$ @ $25 \pm 0.4$ °C          | <u>9.124</u> |
| 2. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)      | <u>Y</u>     |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>Y</u>     |

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  Date: 29 Jul 2010  
VP Oper.  
Title

Rev. 8, 9/19/08

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE