



# CERTIFICATE OF ANALYSIS

## OPA Diluent for Post-Column Carbamate Analysis

CB910

**Product:** OPA Diluent for Post Column Carbamate Analysis

**Part Number:** CB910

**Lot Number:** 011002

**Test Start Date:** 06 Oct 2010

**Expiration Date:** 05 Oct 2012

**Primary Application:** For preparation of post-column fluorescence reagent for carbamate pesticide analysis.

**Acceptance Criteria:**

- 1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C 9.123
- 2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) Y
- 3. Twelve appropriate peaks in chromatogram (Y/N) Y

**Calibration Standard:** Carbamate Test Standard, PN 1700-0063, LN 908045

**Storage Conditions:** Dry at 8 to 25 °C

Analysis Certified by: \_\_\_\_\_

Jim Murphy  
VP Oper  
Title

Date: \_\_\_\_\_

20 Oct 2010

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE