



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-Column Carbamate Analysis

CB910

Product: OPA Diluent for Post Column Carbamate Analysis

Part Number: CB910

Lot Number: 101020

Test Start Date: 20 JAN 2011

Expiration Date: 19 JAN 2013

Primary Application: For preparation of post-column fluorescence reagent for carbamate pesticide analysis.

Acceptance Criteria:

- | | |
|---|--------------|
| 1. pH = 9.10 ± 0.1 @ 25 ± 0.4 °C | <u>9.174</u> |
| 2. Baseline noise (0 ± 25 mvolts) (Y/N) | <u>Y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>Y</u> |

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  **Date:** 15 FEB 2011

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE