



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-Column Carbamate Pesticides Analysis

CB910

Product: OPA Diluent for Post Column analysis of Carbamate pesticides

Part Number: CB910

Lot Number: 10 2036

Manufacture Date: 25 Feb 2011

Expiration Date: 24 Feb 2013

Primary Application: For preparation of post-column fluorescence reagent for Carbamate analysis.

Acceptance Criteria:

Actual Values

- | | |
|---|-------|
| 1. pH = 9.10 ± 0.1 @ 25 ± 0.4 °C | 9.140 |
| 2. Baseline noise (0 ± 25 mvolts) (Y/N) | Y |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | Y |

Sample: Carbamate Test Mixture, **PN:** 1700-0063, **LN:** 908045

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by:

Date: 02 MAR 2011

Title: VP OPR

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 9, 2/16/2011