



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Carbamate Pesticides Analysis

CB910

Product: OPA Diluent for Post-column analysis of Carbamate pesticides

Part Number: CB910

Lot Number: 103006

Production Date: 01 MAR 2011

Expiration Date: 01 MAR 2013

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

1. pH = 9.10 ± 0.1 @ 25 ± 0.4 °C (Y/N) 4
2. Baseline noise (0 ± 25 mvolts) (Y/N) 4
3. Twelve appropriate peaks in chromatogram (Y/N) 4

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8 °C to 25 °C

Analysis Certified by: *[Signature]* **Date:** 14 MAR 2011

VP OPR
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE