



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Carbamate Pesticides Analysis

CB910

Product: OPA Diluent for Post-column analysis of Carbamate pesticides

Part Number: CB910

Lot Number: 104039

Production Date: 27 Apr 2011

Expiration Date: 26 Apr 2013


Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

- 1. pH = 9.10 ± 0.1 @ 25 ± 0.4 °C (Y/N) Y
- 2. Baseline noise (0 ± 25 mvolts) (Y/N) Y
- 3. Twelve appropriate peaks in chromatogram (Y/N) Y

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8 °C to 25 °C

Analysis Certified by:  **Date:** 29 Apr 2011
VP OPR
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE