



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Carbamate Pesticides Analysis

CB910

Product: OPA Diluent for Post-column analysis of Carbamate pesticides

Part Number: CB910

Lot Number: 108024

Production Date: 24 Aug 2011

Expiration Date: 22 Aug 2013

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

1. pH = 9.10 ± 0.1 @ 25 ± 0.4 °C (Y/N) Y
2. Baseline noise (0 ± 25 mvolts) (Y/N) Y
3. Twelve appropriate peaks in chromatogram (Y/N) Y

Sample: Carbamate Test Mixture, PN: 1700-0063

Storage Conditions: Dry at 8 °C to 25 °C

Analysis Certified by: *Jim Murphy* **Date:** 20 Sep 2011
VP Opn
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE