



## CERTIFICATE OF ANALYSIS

### OPA Diluent for Post-column Carbamate Pesticides Analysis

**CB910**

**Product:** OPA Diluent for Post-column analysis of Carbamate pesticides

**Part Number:** CB910

**Lot Number:** 111001

**Production Date:** 02 Nov 2011

**Expiration Date:** 01 Nov 2013

**Primary Application:** Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

1. pH =  $9.10 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N) 4
2. Baseline noise ( $0 \pm 25$  mvolts) (Y/N) 4
3. Twelve appropriate peaks in chromatogram (Y/N) 4

**Sample:** Carbamate Test Mixture, PN: 1700-0063

**Storage Conditions:** Dry at 8 °C to 25 °C

**Analysis Certified by:** *Steve Murphy* **Date:** 08 Nov 2011  
VP Operations  
Title

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE