



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Carbamate Pesticides Analysis

CB910

Product: OPA Diluent for Post-column analysis of Carbamate pesticides

Part Number: CB910

Lot Number: 202026

Production Date: 22 Feb 2012

Expiration Date: 21 Feb 2014

Primary Application: Post-column analysis of Carbamate pesticides.

Acceptance Criteria

1. pH = 9.10 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Baseline noise (0 ± 25 mvolts) (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

4
4
4

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8 °C to 25 °C

Analysis Certified by:

Jim Murphy
VP Oper
Title

Date:

28 Feb 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE