



## CERTIFICATE OF ANALYSIS

### OPA Diluent for Post-column Carbamate Pesticides Analysis

#### CB910

**Product:** OPA Diluent for Post-column analysis of Carbamate pesticides

**Part Number:** CB910

**Lot Number:** 2508011

**Production Date:** 06 Aug 2025

**Expiration Date:** 06 Aug 2027

**Primary Application:** Post-column analysis of Carbamate pesticides.

#### Acceptance Criteria

- |                                                   |          |
|---------------------------------------------------|----------|
| 1. pH = $9.10 \pm 0.1$ @ $25 \pm 0.4$ °C (Y/N)    | <u>Y</u> |
| 2. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)      | <u>Y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>Y</u> |

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063

**Storage Conditions:** Dry at 8 °C to 25 °C

|                               |                     |              |                    |
|-------------------------------|---------------------|--------------|--------------------|
| <b>Analysis Certified by:</b> | <u>David Mazawa</u> | <b>Date:</b> | <u>12 Aug 2025</u> |
| <b>Title:</b>                 | <u>QC Manager</u>   |              |                    |