



CERTIFICATE OF ANALYSIS
OPA Diluent for Glyphosate Analysis
GA104

Product: OPA Diluent for Glyphosate Analysis
Part Number: GA104
Lot Number: 008001
Test Start Date: 02 Aug 2010
Expiration Date: 02 Aug 2012


Primary Application: For preparation of post-column fluorescence reagent for glyphosate herbicide analysis.

Acceptance Criteria:

1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C 10.530
2. Glyphosate and AMPA peaks are visible and in order(Y/N) Y
3. AMPA and Glyphosate area meet specifications (Y/N) Y
NOTE: Area ratios are affected by the hypochlorite added to GA116
4. Baseline noise (0 ± 25 mvolts) (Y/N) Y

Calibration Standard: Glyphosate Test Standard, **PN** 1700-0080, **LN** 912031

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  **Date:** 04 Aug 2010
UP Opr
Title

Rev. 8, 09/25/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE