



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Glyphosate and AMPA Analysis

GA104

Product: OPA Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA104

Lot Number: 104038

Production Date: 02 MAY 2011

Expiration Date: 02 MAY 2013

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) y

2. Glyphosate and AMPA peaks are visible and in order(Y/N) y

3. Baseline noise (0 ± 25 mvolts) (Y/N) y

Sample: Glyphosate Test Mixture, **PN** 1700-0080

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy

Date: 5 MAY 2011

 V.P. OPR
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE