



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Glyphosate and AMPA Analysis

GA104

Product: OPA Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA104

Lot Number: 106002

Production Date: 02 JUN 2011

Expiration Date: 01 JUN 2013

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) 4
2. Glyphosate and AMPA peaks are visible and in order(Y/N) 4
3. Baseline noise (0 ± 25 mvolts) (Y/N) 4

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** 06 Jun 2011

UP OPS
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE