PICKERING
LABORATORIES

# CERTIFICATE OF ANALYSIS

## OPA Diluent for Post-column Glyphosate and AMPA Analysis

## GA104

**Product:** OPA Diluent for Post-column Glyphosate and AMPA analysis
**Part Number:** GA104
**Lot Number:** 108 011
**Production Date:** 26 Aug 2011
**Expiration Date:** 26 Aug 2013

**Primary Application:** Post-column analysis of Glyphosate herbicide.

**Acceptance Criteria:**

1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) Y
2. Glyphosate and AMPA peaks are visible and in order(Y/N) Y
3. Baseline noise (0 ± 25 mvolts) (Y/N) Y

**Sample:** Glyphosate Test Mixture, **PN** 1700-0080

**Storage Conditions: Dry at 8 to 25 °C**

**Analysis Certified by:** [signature] **Date:** 20 Sep 2011
VP Opr
**Title**

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev.A, 04/18/2011 P:\Documentation\Chemical Documentation\GA104

1280 Space Park Way · Mountain View CA 94043 · USA
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · www.pickeringlabs.com · support@pickeringlabs.com