



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Glyphosate and AMPA Analysis

GA104

Product: OPA Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA104

Lot Number: 111021

Production Date: 18 Nov 2011

Expiration Date: 17 Nov 2013

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)

4

2. Glyphosate and AMPA peaks are visible and in order(Y/N)

4


3. Baseline noise (0 ± 25 mvolts) (Y/N)

4

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:


Title

Date: 23 Nov 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE