



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Glyphosate and AMPA Analysis

GA104

Product: OPA Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA104

Lot Number: 203016

Production Date: 12 MAR 2012

Expiration Date: 12 MAR 2014

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)
2. Glyphosate and AMPA peaks are visible and in order(Y/N)
3. Baseline noise (0 ± 25 mvolts) (Y/N)

Y
Y
Y

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

Jim Murphy
VP OPER
Title

Date:

19 MAR 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev.A, 04/18/2011

P:\Documentation\Chemical Documentation\GA104