



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Glyphosate and AMPA Analysis

GA104

Product: OPA Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA104

Lot Number: 2508030

Production Date: 18 Aug 2025

Expiration Date: 18 Aug 2027

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) | <u>Y</u> |
| 2. Glyphosate and AMPA peaks are visible and in order(Y/N) | <u>Y</u> |
| 3. Baseline noise (0 ± 25 mvolts) (Y/N) | <u>Y</u> |

Sample: Glyphosate Test Mixture, **PN:** 1700-0080

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:	<u>David Mazawa</u>	Date:	<u>02 Sep 2025</u>
Title:	<u>QC Manager</u>		