



## CERTIFICATE OF ANALYSIS

### OPA Diluent for Post-column Glyphosate and AMPA Analysis

**GA104**

**Product:** OPA Diluent for Post-column Glyphosate and AMPA analysis

**Part Number:** GA104

**Lot Number:** 711032

**Production Date:** 20 Nov 2017

**Expiration Date:** 20 Nov 2019

**Primary Application:** Post-column analysis of Glyphosate herbicide.

**Acceptance Criteria:**

- |  |          |
|--|----------|
| 1. pH = $10.5 \pm 0.1$ @ $25 \pm 0.4$ °C (Y/N)             | <u>y</u> |
| 2. Glyphosate and AMPA peaks are visible and in order(Y/N) | <u>y</u> |
| 3. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)               | <u>y</u> |

**Sample:** Glyphosate Test Mixture, **PN** 1700-0080

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** December 4, 2017  
President  
Title