



CERTIFICATE OF ANALYSIS

Hypochlorite Diluent for Glyphosate Analysis

GA116

Product: Hypochlorite Diluent for Glyphosate Analysis

Part Number: GA116

Lot Number: 006026

Test Start Date: 12 Jun 2010

Expiration Date: 18 Jun 2012

Primary Application: For preparation of post-column oxidizing reagent for Glyphosate herbicide analysis.

Acceptance Criteria:

1. pH = $11.65 \pm 0.1 @ 25 \pm 0.4 ^\circ\text{C}$

2. Glyphosate & AMPA peaks are visible and in order (Y/N)

3. AMPA and Glyphosate area meet specifications (Y/N)

NOTE: Area ratios are affected by the hypochlorite added to GA116

4. Baseline noise (0 ± 25 mvolts) (Y/N)

Actual Values

11.648

Y

Y

Y


Y

Calibration Standard: Glyphosate Test Standard, PN 1700-0080, LN

912031

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____


VP Ops
Title

Date: _____

23 Jun 2010

Rev. 9, 09/25/08

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE