



CERTIFICATE OF ANALYSIS

Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

GA116

Product: Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA116

Lot Number: 103014

Production Date: 07 MAR 2011

Expiration Date: 07 MAR 2013

Primary Application: For preparation of post-column oxidizing reagent for Glyphosate herbicide analysis.

Acceptance Criteria:

1. pH = 11.65 ± 0.1 @ 25 ± 0.4 °C (Y/N)

Y

2. Glyphosate & AMPA peaks are visible and in order (Y/N)

Y

4. Baseline noise (0 ± 25 mvolts) (Y/N)

Y

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: _____

Jim [Signature]
VP OPR

Title

Date: _____

14 MAR 2011

Rev. 10, 04/06/2011

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE