



CERTIFICATE OF ANALYSIS

Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

GA116

Product: Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA116

Lot Number: 108030

Production Date: 26 AUG 2011

Expiration Date: 25 AUG 2013

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

1. pH = 11.65 ± 0.1 @ 25 ± 0.4 °C (Y/N) y

2. Glyphosate & AMPA peaks are visible and in order (Y/N) y

4. Baseline noise (0 ± 25 mvolts) (Y/N) y

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: 

Date: 20 Sep 2011

Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. A, 04/18/2011

P:\Documentation\Chemical Documentation\GA116