



CERTIFICATE OF ANALYSIS

Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

GA116

Product: Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA116

Lot Number: 109013

Production Date: 19 Sep 2011

Expiration Date: 12 Sep 2013

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

1. pH = 11.65 ± 0.1 @ 25 ± 0.4 °C (Y/N) Y

2. Glyphosate & AMPA peaks are visible and in order (Y/N) Y

4. Baseline noise (0 ± 25 mvolts) (Y/N) Y

Sample: Glyphosate Test Mixture, PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: 

Date: 21 Sep 2011

V.P. O'R
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE