



## CERTIFICATE OF ANALYSIS

### Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

**GA116**

**Product:** Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis

**Part Number:** GA116

**Lot Number:** 202018

**Production Date:** 29 Feb 2012

**Expiration Date:** 28 Feb 2014

**Primary Application:** Post-column analysis of Glyphosate herbicide.

**Acceptance Criteria:**

1. pH =  $11.65 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)
2. Glyphosate & AMPA peaks are visible and in order (Y/N)
3. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)

4  
4  
4

**Sample:** Glyphosate Test Mixture, PN 1700-0080

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:**

*Jim Myers*  
VP Corp  
Title

**Date:**

05 Mar 2011

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE