



## CERTIFICATE OF ANALYSIS

### Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

**GA116**

**Product:** Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis

**Part Number:** GA116

**Lot Number:** 204003

**Production Date:** 02 Apr 2012

**Expiration Date:** 02 Apr 2014

**Primary Application:** Post-column analysis of Glyphosate herbicide.

**Acceptance Criteria:** 1.2.11.1.2.3

1. pH =  $11.65 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)
2. Glyphosate & AMPA peaks are visible and in order (Y/N)
3. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)

Y  
Y  
Y

**Sample:** Glyphosate Test Mixture, PN 1700-0080

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:**

[Signature]  
Title

**Date:**

06 Apr 2012

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. A, 04/18/2011

P:\Documentation\Chemical Documentation\GA116