



CERTIFICATE OF ANALYSIS

Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

GA116

Product: Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis

Part Number: GA116

Lot Number: 2308017

Production Date: 09 Aug 2023

Expiration Date: 09 Aug 2025

Primary Application: Post-column analysis of Glyphosate herbicide.

Acceptance Criteria:

- | | |
|---|----------|
| 1. pH = 11.65 ± 0.1 @ 25 ± 0.4 °C (Y/N) | <u>Y</u> |
| 2. Glyphosate & AMPA peaks are visible and in order (Y/N) | <u>Y</u> |
| 3. Baseline noise (0 ± 25 mvolts) (Y/N) | <u>Y</u> |

Sample: Glyphosate Test Mixture, **PN** 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: David Mazawa **Date:** 11 Aug 2023
Title: Technical Support Chemist