



## CERTIFICATE OF ANALYSIS

### Hypochlorite Diluent for Post-column Glyphosate and AMPA Analysis

**GA116**

**Product:** Hypochlorite Diluent for Post-column Glyphosate and AMPA analysis

**Part Number:** GA116

**Lot Number:** 709023

**Production Date:** 21 Sep 2017

**Expiration Date:** 21 Sep 2019

**Primary Application:** Post-column analysis of Glyphosate herbicide.

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. pH = $11.65 \pm 0.1$ @ $25 \pm 0.4$ °C (Y/N)           | <u>y</u> |
| 2. Glyphosate & AMPA peaks are visible and in order (Y/N) | <u>y</u> |
| 3. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)              | <u>y</u> |

**Sample:** Glyphosate Test Mixture, **PN** 1700-0080

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** 28 Sept 2017  
President  
Title