



## CERTIFICATE OF ANALYSIS

### Modified K600 Potassium Eluant For Invitrogen Life Technologies IVGN-K600

**Product:** Modified K600 Potassium Eluent

**Part Number:** IVGN-K600

**Lot Number:** 708062

**Production Date:** 31 Aug 2017

**Expiration Date:** 31 Aug 2019

**Primary Application:** Custom K600 Eluent for Invitrogen Life Technologies

**Acceptance Criteria:**

1. pH =  $6.30 \pm 0.05$  @  $25 \pm 0.4$  °C (Y/N)

y

**Storage Conditions:** Dry at 8 to 25°C

**Analysis Certified by:** Jim Murphy

President

Title

**Date:** September 5, 2017