



CERTIFICATE OF ANALYSIS

Potassium Eluant For Biogenic Amines Analysis

K130

Product: Potassium Eluant for Biogenic Amines Analysis

Part Number: K130

Lot Number: 2511008

Production Date: 06 Nov 2025

Expiration Date: 06 Nov 2027

Primary Application: Eluant for ALKION Cation-exchange chromatography of polyamines.

Acceptance Criteria:

1. pH $\geq$ 12.86 @ 25 $\pm$ 0.4 °C (Y/N) Y

Storage Conditions: Dry at 8 to 25°C

Analysis Certified by: David Mazawa **Date:** 11 Nov 2025
Title: QC Manager