

## CERTIFICATE OF ANALYSIS

### K200 Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

**Product:** K200 Eluant for ion-exchange chromatography of glyphosate and AMPA  
**Part Number:** K200  
**Lot Number:** 006013  
**Test Start Date:** 10 JUN 2010  
**Expiration Date:** 10 JUN 2012

**Primary Application:** Eluant for ion-exchange chromatography of glyphosate and AMPA


**Acceptance Criteria:**

1. pH =  $2.00 \pm 0.05$  at  $25 \pm 0.4$  °C
  2. Elution times meet specifications
    - Glyphosate (Y/N)
    - AMPA (Y/N)
  3. AMPA and Glyphosate area meet specifications (Y/N)
- NOTE:** Area ratios are affected by the hypochlorite added to GA116
4. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)

**Actual Values**

2.050  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:**  **Date:** 21 JUN 2010

VP  
 \_\_\_\_\_  
 Title