



CERTIFICATE OF ANALYSIS

K200 Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

Product: K200 Eluant for ion-exchange chromatography of glyphosate and AMPA
Part Number: K200
Lot Number: 007037
Test Start Date: 29 Jul 2010
Expiration Date: 28 Jul 2012

Primary Application: Eluant for ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

- pH = 2.00 ± 0.05 at 25 ± 0.4 °C
- Elution times meet specifications
 - Glyphosate (Y/N)
 - AMPA (Y/N)
- AMPA and Glyphosate area meet specifications (Y/N)
NOTE: Area ratios are affected by the hypochlorite added to GA116
- Baseline noise (0 ± 25 mvolts) (Y/N)

Actual Values

2.027
Y
Y
Y
Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Murphy
VP O&R
Title

Date: 04 Aug 2010

Rev. 6, 8/11/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE