



# CERTIFICATE OF ANALYSIS

## K200 Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

Product: K200 Eluant for ion-exchange chromatography of glyphosate and AMPA  
Part Number: K200  
Lot Number: 009029  
Test Start Date: 16 Sep 2010  
Expiration Date: 16 Sep 2012

Primary Application: Eluant for Ion-exchange chromatography of glyphosate and AMPA

### Acceptance Criteria:

1. pH =  $2.00 \pm 0.05$  at  $25 \pm 0.4$  °C
  2. Elution times meet specifications
    - Glyphosate (Y/N)
    - AMPA (Y/N)
  3. AMPA and Glyphosate area meet specifications (Y/N)
  4. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)
- NOTE:** Area ratios are affected by the hypochlorite added to GA116

### Actual Values

<u>2.003</u>
<u>Y</u>
<u>Y</u>
<u>Y</u>
<u>Y</u>

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: [Signature]  
Title: VIP

Date: 22 Sep 2010

Rev. 6, 8/11/08

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE