



CERTIFICATE OF ANALYSIS

K200 Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

Product: K200 Eluant for ion-exchange chromatography of glyphosate and AMPA
Part Number: K200
Lot Number: 010003
Test Start Date: 06 Oct 2010
Expiration Date: 06 Oct 2012

Primary Application: Eluant for ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

1. pH = 2.00 ± 0.05 at 25 ± 0.4 °C
2. Elution times meet specifications
 - Glyphosate (Y/N)
 - AMPA (Y/N)
3. AMPA and Glyphosate area meet specifications (Y/N)
NOTE: Area ratios are affected by the hypochlorite added to GA116
4. Baseline noise (0 ± 25 mvolts) (Y/N)

Actual Values

2.013
Y
Y
Y
Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Murphy
VP Ops

Title

Date: 20 Oct 2010

Rev. 6, 8/11/08

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE