



# CERTIFICATE OF ANALYSIS

## K200 Eluant for Ion-exchange Chromatography of Glyphosate and AMPA

K200

**Product:** K200 Eluant for ion-exchange chromatography of glyphosate and AMPA

**Part Number:** K200

**Lot Number:** 101001

**Test Start Date:** 05 JAN 2011

**Expiration Date:** 04 JAN 2013

**Primary Application:** Eluant for ion-exchange chromatography of glyphosate and AMPA

### Acceptance Criteria:

1. pH =  $2.00 \pm 0.05$  at  $25 \pm 0.4$  °C

2. Elution times meet specifications

- Glyphosate (Y/N)

- AMPA (Y/N)

3. AMPA and Glyphosate area meet specifications (Y/N)

**NOTE:** Area ratios are affected by the hypochlorite added to GA116

4. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)

### Actual Values

1.983

Y

Y

Y

Y

**Storage Conditions:** Dry at 8 to 25 °C

Analysis Certified by: \_\_\_\_\_

*Jim Murphy*  
VP Ops

Title

Date: 18 JAN 2011

Rev. 6, 8/11/08

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE