

CERTIFICATE OF ANALYSIS

K200 Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

Product: K200 Eluant for ion-exchange chromatography of glyphosate and AMPA
Part Number: K200
Lot Number: 102010
Test Start Date: 02/09/11
Expiration Date: 09 Feb 2013


Primary Application: Eluant for Ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

1. pH = 2.00 ± 0.05 at 25 ± 0.4 °C
2. Elution times meet specifications
 - Glyphosate (Y/N)
 - AMPA (Y/N)
3. AMPA and Glyphosate area meet specifications (Y/N)
 NOTE: Area ratios are affected by the hypochlorite added to GA116
4. Baseline noise (0 ± 25 mvolts) (Y/N)

1.976
Actual Values
1.976

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  **Date:** 15 Feb 2011

 VP OPR

 Title

ACCEPTED
 PICKERING LABORATORIES, INC.
 QUALITY ASSURANCE

Rev. 6, 8/11/08