



CERTIFICATE OF ANALYSIS

Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

Product: K200 Eluant for ion-exchange chromatography of glyphosate and AMPA

Part Number: K200

Lot Number: 103018

Production Date: 09 MAR 2011

Expiration Date: 09 MAR 2013

Primary Application: Eluant for ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

1. pH = 2.00 ± 0.05 at 25 ± 0.4 °C (Y/N) 4

2. Elution times meet specifications

• Glyphosate (Y/N) 4

• AMPA (Y/N) 4

3. AMPA and Glyphosate area meet specifications (Y/N) 4

NOTE: Area ratios are affected by the hypochlorite added to GA116

4. Baseline noise (0 ± 25 mvolts) (Y/N) 4

Sample: Glyphosate Test Mixture: PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: Jim Murphy

Date: 21 MAR 2011

VP OPR
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 7, 3/25/2011