

CERTIFICATE OF ANALYSIS

Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

Product: Eluant for ion-exchange chromatography of glyphosate and AMPA

Part Number: K200

Lot Number: 105009

Production Date: 10 MAY 2011

Expiration Date: 10 MAY 2013

Primary Application: Eluant for Ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

1. pH = 2.00 ± 0.05 at 25 ± 0.4 °C (Y/N)
2. Elution times meet specifications
 - Glyphosate (Y/N)
 - AMPA (Y/N)
3. AMPA and Glyphosate area meet specifications (Y/N)
4. Baseline noise (0 ± 25 mvolts) (Y/N)

Y
Y
Y
Y
Y

Sample: Glyphosate Test Mixture: PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: Jim Murphy

Date: 18 MAY 2011

VP OPR
Title

ACCEPTED

PICKERING LABORATORIES, INC
QUALITY ASSURANCE