



## CERTIFICATE OF ANALYSIS

### Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

**Product:** Eluant for ion-exchange chromatography of glyphosate and AMPA

**Part Number:** K200

**Lot Number:** 107015

**Production Date:** 14 Jul 2011

**Expiration Date:** 12 Jul 2013

**Primary Application:** Eluant for ion-exchange chromatography of glyphosate and AMPA

**Acceptance Criteria:**

1. pH =  $2.00 \pm 0.05$  at  $25 \pm 0.4$  °C (Y/N)
2. Elution times meet specifications
  - Glyphosate (Y/N)
  - AMPA (Y/N)
3. AMPA and Glyphosate area meet specifications (Y/N)
4. Baseline noise ( $0 \pm 25$  mvolts) (Y/N)

4  
4  
4  
4  
4

**Sample:** Glyphosate Test Mixture: PN 1700-0080

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:** \_\_\_\_\_

*[Signature]*  
VP OPR

Title

**Date:** 15 Jul 2011

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE