



CERTIFICATE OF ANALYSIS

Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

Product: Eluant for ion-exchange chromatography of glyphosate and AMPA

Part Number: K200

Lot Number: 107036

Production Date: 29 Jul 2011

Expiration Date: 29 Jul 2013

Primary Application: Eluant for ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

1. pH = 2.00 ± 0.05 at 25 ± 0.4 °C (Y/N)
2. Elution times meet specifications
 - Glyphosate (Y/N)
 - AMPA (Y/N)
3. AMPA and Glyphosate area meet specifications (Y/N)
4. Baseline noise (0 ± 25 mvolts) (Y/N)

y

y

y

y

y

Sample: Glyphosate Test Mixture: PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: _____

Date: 10 Aug 2011

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. A, 04/29/2011

P:\Documentation\Chemical Documentation\K200