



CERTIFICATE OF ANALYSIS

Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

Product: Eluant for ion-exchange chromatography of glyphosate and AMPA

Part Number: K200

Lot Number: 202007

Production Date: 06 Feb 2012

Expiration Date: 03 Feb 2014

Primary Application: Eluant for ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

1. pH = 2.00 ± 0.05 at 25 ± 0.4 °C (Y/N)
2. Elution times meet specifications
 - Glyphosate (Y/N)
 - AMPA (Y/N)
3. AMPA and Glyphosate area meet specifications (Y/N)
4. Baseline noise (0 ± 25 mvolts) (Y/N)

Y
Y
Y
Y
Y

Sample: Glyphosate Test Mixture: PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: *[Signature]*

VP Support
Title

Date: 28 Feb 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE