



CERTIFICATE OF ANALYSIS

Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

Product: Eluant for ion-exchange chromatography of glyphosate and AMPA

Part Number: K200

Lot Number: 2508019

Production Date: 13 Aug 2025

Expiration Date: 13 Aug 2027

Primary Application: Eluant for ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

1. pH = 2.00 ± 0.05 at 25 ± 0.4 °C (Y/N)	<u>Y</u>
2. Elution times meet specifications	
• Glyphosate (Y/N)	<u>Y</u>
• AMPA (Y/N)	<u>Y</u>
3. AMPA and Glyphosate area meet specifications (Y/N)	<u>Y</u>
4. Baseline noise (0 ± 25 mvolts) (Y/N)	<u>Y</u>

Sample: Glyphosate Test Mixture: PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: David Mazawa **Date:** 18 Aug 2025
Title: QC Manager