CERTIFICATE OF ANALYSIS

Eluant for Ion-exchange Chromatography of Glyphosate and AMPA
K200

Product: Eluant for ion-exchange chromatography of glyphosate and AMPA
Part Number: K200
Lot Number: 306006
Production Date: 06 Jun 2013
Expiration Date: 06 Jun 2015

Primary Application: Eluant for Ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

1. pH = 2.00 ± 0.05 at 25 ± 0.4 °C (Y/N) **y**
2. Elution times meet specifications
   - Glyphosate (Y/N) **y**
   - AMPA (Y/N) **y**
3. AMPA and Glyphosate area meet specifications (Y/N) **y**
4. Baseline noise (0 ± 25 mvolts) (Y/N) **y**

Sample: Glyphosate Test Mixture: PN 1700-0080

Storage Conditions: Dry at 8°C to 25 °C

Analysis Certified by: Jim Murphy **Date:** 13 Jun 2013

President
Title

Rev. A, 04/29/2011 **P:\Documentation\Chemical Documentation\K200**

1280 Space Park Way · Mountain View CA 94043 · USA
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · www.pickeringlabs.com · support@pickeringlabs.com