



## CERTIFICATE OF ANALYSIS

### Eluant for Ion-exchange Chromatography of Glyphosate and AMPA K200

**Product:** Eluant for ion-exchange chromatography of glyphosate and AMPA

**Part Number:** K200

**Lot Number:** 501018

**Production Date:** 14 Jan 2015

**Expiration Date:** 14 Jan 2017

**Primary Application:** Eluant for ion-exchange chromatography of glyphosate and AMPA

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. pH = $2.00 \pm 0.05$ at $25 \pm 0.4$ °C (Y/N)      | <u>y</u> |
| 2. Elution times meet specifications                  |          |
| • Glyphosate (Y/N)                                    | <u>y</u> |
| • AMPA (Y/N)  | <u>y</u> |
| 3. AMPA and Glyphosate area meet specifications (Y/N) | <u>y</u> |
| 4. Baseline noise ( $0 \pm 25$ mvolts) (Y/N)          | <u>y</u> |

**Sample:** Glyphosate Test Mixture: PN 1700-0080

**Storage Conditions:** Dry at 8°C to 25 °C

**Analysis Certified by:** Jim Murphy      **Date:** 20 Jan 2015  
President  
Title