



# CERTIFICATE OF ANALYSIS

## Potassium Eluant For Biogenic Amines Analysis

**K563**

**Product:** K563 Potassium Eluant for Biogenic Amines Analysis

**Part Number:** K563

**Lot Number:** 006032

**Test Start Date:** 22 Jun 2010

**Primary Application:** Eluant for Ion-exchange chromatography of glyphosate and AMPA

**Acceptance Criteria:**

**Values**


1. pH =  $5.63 \pm 0.10$  @  $25 \pm 0.4$  °C

5.531

2. Peak resolution meets specification (y/n):

y

**Analysis Certified by:**



**Date:**

23 Jun 2010

Title

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE