



CERTIFICATE OF ANALYSIS

Potassium Eluant For Biogenic Amines Analysis

K563

Product: K563 Potassium Eluant for Biogenic Amines Analysis

Part Number: K563

Lot Number: 103011

Test Start Date: 03 MAR 2011

Primary Application: Eluant for Ion-exchange chromatography of glyphosate and AMPA

Acceptance Criteria:

Values

1. pH = 5.63 ± 0.10 @ 25 ± 0.4 °C

5.642

2. Peak resolution meets specification (y/n):

y

Analysis Certified by:

Jim Murphy
VP OPR
Title

Date:

14 MAR 2011

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE